

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042120\
 Data File : VX015799.D
 Acq On : 21 Apr 2020 17:02
 Operator : JC/SP
 Sample : L2327-18
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-4-S

Quant Time: Apr 22 06:16:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1053955	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1617366	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1522714	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	879611	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	594731	40.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.02%	
35) Dibromofluoromethane	5.48	113	484818	46.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.78%	
50) Toluene-d8	8.70	98	1922154	50.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.00%	
62) 4-Bromofluorobenzene	11.12	95	804299	51.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.98%	
Target Compounds						
16) Acetone	2.42	43	163551	25.610	ug/l	95
17) Carbon Disulfide	2.56	76	85640	2.821	ug/l	96
20) Methylene Chloride	2.84	84	58041	4.866	ug/l	95
43) Isopropyl Acetate	6.42	43	367891	11.231	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042120\
Data File : VX015799.D
Acq On : 21 Apr 2020 17:02
Operator : JC/SP
Sample : L2327-18
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-4-S

Quant Time: Apr 22 06:16:40 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

