

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042120\
 Data File : VX015801.D
 Acq On : 21 Apr 2020 17:47
 Operator : JC/SP
 Sample : L2327-24
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-6-S

Quant Time: Apr 22 06:20:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1040050	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1610433	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1538676	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	863652	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	587762	40.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.14%	
35) Dibromofluoromethane	5.47	113	476298	46.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.54%	
50) Toluene-d8	8.70	98	1896936	50.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.10%	
62) 4-Bromofluorobenzene	11.12	95	806534	52.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.72%	
Target Compounds						
16) Acetone	2.42	43	148223	23.520	ug/l	93
17) Carbon Disulfide	2.56	76	83713	2.795	ug/l	98
20) Methylene Chloride	2.84	84	72693	6.175	ug/l #	92
43) Isopropyl Acetate	6.42	43	353756	10.846	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042120\
Data File : VX015801.D
Acq On : 21 Apr 2020 17:47
Operator : JC/SP
Sample : L2327-24
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-6-S

Quant Time: Apr 22 06:20:37 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

