

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042120\
 Data File : VX015805.D
 Acq On : 21 Apr 2020 19:17
 Operator : JC/SP
 Sample : L2323-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-19-COMP

Quant Time: Apr 22 06:29:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1021567	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1539266	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1473504	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	848010	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	564932	39.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.42%	
35) Dibromofluoromethane	5.47	113	447953	45.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.06%	
50) Toluene-d8	8.70	98	1881989	51.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.90%	
62) 4-Bromofluorobenzene	11.12	95	786062	53.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.78%	
Target Compounds						
16) Acetone	2.42	43	155525	25.125	ug/l	98
17) Carbon Disulfide	2.56	76	45815	1.557	ug/l #	88
20) Methylene Chloride	2.84	84	41844	3.619	ug/l	93
43) Isopropyl Acetate	6.42	43	334728	10.737	ug/l	98
95) Naphthalene	13.82	128	128274	2.047	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042120\
Data File : VX015805.D
Acq On : 21 Apr 2020 19:17
Operator : JC/SP
Sample : L2323-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-19-COMP

Quant Time: Apr 22 06:29:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

