

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042120\
 Data File : VX015816.D
 Acq On : 22 Apr 2020 00:34
 Operator : JC/SP
 Sample : PB128378TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB128378TB

Quant Time: Apr 22 06:59:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	959980	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1461182	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1429699	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	815579	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	529730	39.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.24%	
35) Dibromofluoromethane	5.48	113	418033	44.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.52%	
50) Toluene-d8	8.70	98	1792236	52.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.24%	
62) 4-Bromofluorobenzene	11.12	95	757097	54.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.34%	
Target Compounds						
16) Acetone	2.42	43	84933	14.601	ug/l	93
17) Carbon Disulfide	2.56	76	33342	1.206	ug/l	100
20) Methylene Chloride	2.84	84	10997	1.012	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042120\
Data File : VX015816.D
Acq On : 22 Apr 2020 00:34
Operator : JC/SP
Sample : PB128378TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB128378TB

Quant Time: Apr 22 06:59:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

