

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042120\
 Data File : VX015817.D
 Acq On : 22 Apr 2020 00:57
 Operator : JC/SP
 Sample : L2352-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SP-LAWNSIDE

Quant Time: Apr 22 07:01:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1008650	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1552241	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1531767	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	849115	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	539988	37.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.92%	
35) Dibromofluoromethane	5.47	113	432140	43.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.10%	
50) Toluene-d8	8.70	98	1860955	50.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.88%	
62) 4-Bromofluorobenzene	11.12	95	787175	53.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.04%	
Target Compounds						
16) Acetone	2.42	43	68638	11.230	ug/l	92
17) Carbon Disulfide	2.56	76	31821	1.095	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042120\
Data File : VX015817.D
Acq On : 22 Apr 2020 00:57
Operator : JC/SP
Sample : L2352-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SP-LAWNSIDE

Quant Time: Apr 22 07:01:48 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

