

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042120\
 Data File : VX015822.D
 Acq On : 22 Apr 2020 02:50
 Operator : JC/SP
 Sample : L2351-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-21A

Quant Time: Apr 22 07:14:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	969624	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1494043	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1452584	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	807148	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	539198	39.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.86%	
35) Dibromofluoromethane	5.48	113	433271	45.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.74%	
50) Toluene-d8	8.70	98	1782645	50.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.40%	
62) 4-Bromofluorobenzene	11.12	95	761130	53.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.52%	
Target Compounds						
16) Acetone	2.42	43	256947	43.733	ug/l	99
20) Methylene Chloride	2.84	84	45957	4.188	ug/l	85
43) Isopropyl Acetate	6.42	43	315296	10.420	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042120\
Data File : VX015822.D
Acq On : 22 Apr 2020 02:50
Operator : JC/SP
Sample : L2351-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-21A

Quant Time: Apr 22 07:14:04 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

