

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042120\
 Data File : VX015823.D
 Acq On : 22 Apr 2020 03:13
 Operator : JC/SP
 Sample : L2353-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VNJ-240

Quant Time: Apr 22 07:16:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	982686	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1493373	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1452738	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	820830	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	533698	38.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.02%	
35) Dibromofluoromethane	5.47	113	426194	44.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.30%	
50) Toluene-d8	8.70	98	1793349	51.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.06%	
62) 4-Bromofluorobenzene	11.12	95	758529	53.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.22%	
Target Compounds						
16) Acetone	2.42	43	102188	17.162	ug/l	97
17) Carbon Disulfide	2.55	76	34399	1.215	ug/l #	91
20) Methylene Chloride	2.84	84	16977	1.526	ug/l #	89
43) Isopropyl Acetate	6.42	43	297738	9.844	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042120\
Data File : VX015823.D
Acq On : 22 Apr 2020 03:13
Operator : JC/SP
Sample : L2353-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VNJ-240

Quant Time: Apr 22 07:16:48 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

