

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042120\
 Data File : VX015825.D
 Acq On : 22 Apr 2020 03:58
 Operator : JC/SP
 Sample : L2357-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-6

Quant Time: Apr 22 07:27:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	957398	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1471071	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1420279	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	810897	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	527708	39.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.16%	
35) Dibromofluoromethane	5.47	113	423286	45.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.02%	
50) Toluene-d8	8.70	98	1785599	51.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.16%	
62) 4-Bromofluorobenzene	11.12	95	753591	53.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.12%	
Target Compounds						
16) Acetone	2.42	43	174830	30.137	ug/l	99
20) Methylene Chloride	2.84	84	43610	4.025	ug/l	89
43) Isopropyl Acetate	6.42	43	301332	10.114	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042120\
Data File : VX015825.D
Acq On : 22 Apr 2020 03:58
Operator : JC/SP
Sample : L2357-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-6

Quant Time: Apr 22 07:27:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

