

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042121\
 Data File : VX021790.D
 Acq On : 21 Apr 2021 14:18
 Operator : JC/MD
 Sample : M2122-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW8-GW-98-100

Quant Time: Apr 22 08:57:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 20 13:46:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	173282	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	295732	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	265736	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	126904	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	99943	50.84	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.68%
35) Dibromofluoromethane	5.464	113	93131	49.24	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.48%
50) Toluene-d8	8.696	98	352829	48.52	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.04%
62) 4-Bromofluorobenzene	11.122	95	127616	46.99	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.98%
Target Compounds						
						Qvalue
12) 1,1-Dichloroethene	2.355	96	1021	0.64	ug/l	92
19) Methyl tert-butyl Ether	3.166	73	3043	0.55	ug/l #	69
24) 1,1-Dichloroethane	3.672	63	6699	2.19	ug/l	97
27) cis-1,2-Dichloroethene	4.574	96	2929	1.40	ug/l	78
30) Chloroform	5.172	83	866	0.26	ug/l #	75
64) Tetrachloroethene	9.317	164	2789	1.34	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042121\
Data File : VX021790.D
Acq On : 21 Apr 2021 14:18
Operator : JC/MD
Sample : M2122-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-98-100

Quant Time: Apr 22 08:57:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 20 13:46:22 2021
Response via : Initial Calibration

