

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042121\
 Data File : VX021800.D
 Acq On : 21 Apr 2021 18:06
 Operator : JC/MD
 Sample : M2114-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-02

Quant Time: Apr 22 09:00:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 20 13:46:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	163442	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	278552	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	246863	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	115323	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	94385	50.91	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.82%
35) Dibromofluoromethane	5.464	113	88202	49.51	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.02%
50) Toluene-d8	8.696	98	329815	48.15	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.30%
62) 4-Bromofluorobenzene	11.122	95	116603	45.59	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.18%
Target Compounds						
16) Acetone	2.422	43	3927	6.16	ug/l	90
20) Methylene Chloride	2.837	84	1154	0.60	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042121\
Data File : VX021800.D
Acq On : 21 Apr 2021 18:06
Operator : JC/MD
Sample : M2114-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB-02

Quant Time: Apr 22 09:00:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 20 13:46:22 2021
Response via : Initial Calibration

