

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042122\
 Data File : VX028236.D
 Acq On : 21 Apr 2022 13:18
 Operator : JC/MD
 Sample : N2468-09ME
 Misc : 5.99g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-15-1-8-GRME

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 05/06/2022
 Supervised By :Semsettin Yesilyurt 05/06/2022

Quant Time: Apr 21 16:42:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	448603	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	735672	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	629844	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	310118	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	251468	42.991	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	85.980%		
35) Dibromofluoromethane	5.391	113	229798	44.541	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	89.080%		
50) Toluene-d8	8.647	98	841448	42.464	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	84.920%		
62) 4-Bromofluorobenzene	11.085	95	286052	37.755	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	75.500%		
Target Compounds						
					Qvalue	
31) Cyclohexane	5.470	56	10019	1.406	ug/l	85
39) Methylcyclohexane	7.379	83	59043	6.953	ug/l	97
65) Chlorobenzene	10.079	112	13677	1.018	ug/l	98
73) Isopropylbenzene	10.963	105	170737	8.010	ug/l	100
80) 1,3,5-Trimethylbenzene	11.457	105	33695m	1.862	ug/l	
84) 1,2,4-Trimethylbenzene	11.756	105	56764	3.156	ug/l	97
85) sec-Butylbenzene	11.896	105	121240	5.439	ug/l #	76
89) n-Butylbenzene	12.335	91	102534	6.494	ug/l #	67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042122\
Data File : VX028236.D
Acq On : 21 Apr 2022 13:18
Operator : JC/MD
Sample : N2468-09ME
Misc : 5.99g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-15-1-8-GRME

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 05/06/2022
Supervised By :Semsettin Yesilyurt 05/06/2022

Quant Time: Apr 21 16:42:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

