

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042122\
 Data File : VX028239.D
 Acq On : 21 Apr 2022 14:27
 Operator : JC/MD
 Sample : N2459-03DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 31573DL

Quant Time: Apr 22 06:12:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	436590	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	744690	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	707141	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	346627	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	301890	53.032	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 106.060%			
35) Dibromofluoromethane	5.391	113	277281	53.094	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 106.180%			
50) Toluene-d8	8.653	98	1011443	50.425	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 100.840%			
62) 4-Bromofluorobenzene	11.079	95	372609	48.583	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 97.160%			
Target Compounds					Qvalue	
16) Acetone	2.379	43	15976	7.836	ug/l	92
25) 2-Butanone	4.562	43	7458	2.212	ug/l #	84
31) Cyclohexane	5.470	56	12477	1.799	ug/l	96
39) Methylcyclohexane	7.391	83	4669	0.543	ug/l #	87
40) Benzene	6.043	78	400146	20.067	ug/l	99
52) Toluene	8.720	92	459102	33.139	ug/l	99
67) Ethyl Benzene	10.195	91	36509	1.444	ug/l	99
68) m/p-Xylenes	10.299	106	282644	27.470	ug/l	97
69) o-Xylene	10.640	106	161785	15.771	ug/l	96
80) 1,3,5-Trimethylbenzene	11.451	105	81200	4.015	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	296445	14.746	ug/l	97
95) Naphthalene	13.780	128	49692	2.165	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042122\
Data File : VX028239.D
Acq On : 21 Apr 2022 14:27
Operator : JC/MD
Sample : N2459-03DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
31573DL

Quant Time: Apr 22 06:12:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

