

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042122\
 Data File : VX028243.D
 Acq On : 21 Apr 2022 16:00
 Operator : JC/MD
 Sample : N2516-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-TB-20220419

Quant Time: Apr 22 06:14:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	397143	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	684302	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	654160	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	309823	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	282934	54.639	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 109.280%	
35) Dibromofluoromethane	5.391	113	258125	53.788	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 107.580%	
50) Toluene-d8	8.653	98	948456	51.457	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 102.920%	
62) 4-Bromofluorobenzene	11.079	95	342570	48.608	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 97.220%	
Target Compounds						
20) Methylene Chloride	2.788	84	1243	0.324	ug/l	Qvalue 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042122\
Data File : VX028243.D
Acq On : 21 Apr 2022 16:00
Operator : JC/MD
Sample : N2516-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-TB-20220419

Quant Time: Apr 22 06:14:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

