

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042122\
 Data File : VX028246.D
 Acq On : 21 Apr 2022 17:09
 Operator : JC/MD
 Sample : N2459-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 31583

Quant Time: Apr 22 06:15:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	382430	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	660290	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	636312	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	303390	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	271874	54.523	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	109.040%
35) Dibromofluoromethane	5.385	113	246142	53.156	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	106.320%
50) Toluene-d8	8.653	98	908403	51.077	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.160%
62) 4-Bromofluorobenzene	11.079	95	335299	49.307	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.620%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	6796	3.083	ug/l #	83
25) 2-Butanone	4.580	43	1301	0.441	ug/l	91
67) Ethyl Benzene	10.195	91	5418	0.238	ug/l	99
68) m/p-Xylenes	10.305	106	3913	0.423	ug/l	99
69) o-Xylene	10.646	106	3265	0.354	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042122\
Data File : VX028246.D
Acq On : 21 Apr 2022 17:09
Operator : JC/MD
Sample : N2459-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
31583

Quant Time: Apr 22 06:15:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

