

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042122\
 Data File : VX028247.D
 Acq On : 21 Apr 2022 17:33
 Operator : JC/MD
 Sample : N2459-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 31584

Quant Time: Apr 22 06:16:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	382783	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	663620	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	636346	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	302849	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	281660	56.433	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	112.860%
35) Dibromofluoromethane	5.391	113	255149	54.825	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	109.640%
50) Toluene-d8	8.653	98	928350	51.936	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.880%
62) 4-Bromofluorobenzene	11.079	95	342493	50.112	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.220%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	6044	2.583	ug/l	89
25) 2-Butanone	4.574	43	1473	0.498	ug/l #	75
69) o-Xylene	10.640	106	2211	0.240	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042122\
Data File : VX028247.D
Acq On : 21 Apr 2022 17:33
Operator : JC/MD
Sample : N2459-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
31584

Quant Time: Apr 22 06:16:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

