

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042122\
 Data File : VX028252.D
 Acq On : 21 Apr 2022 19:29
 Operator : JC/MD
 Sample : N2459-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 31580

Quant Time: Apr 22 06:18:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	394437	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	677979	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	636874	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	300229	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	279372	54.321	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 108.640%	
35) Dibromofluoromethane	5.391	113	254010	53.424	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 106.840%	
50) Toluene-d8	8.653	98	925724	50.693	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 101.380%	
62) 4-Bromofluorobenzene	11.079	95	335511	48.051	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 96.100%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	6431	2.713	ug/l #	79
40) Benzene	6.043	78	3936	0.217	ug/l #	90
52) Toluene	8.726	92	11178	0.886	ug/l	99
67) Ethyl Benzene	10.201	91	8819	0.387	ug/l	97
68) m/p-Xylenes	10.305	106	13039	1.407	ug/l	96
69) o-Xylene	10.646	106	8267	0.895	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	4687	0.269	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042122\
Data File : VX028252.D
Acq On : 21 Apr 2022 19:29
Operator : JC/MD
Sample : N2459-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
31580

Quant Time: Apr 22 06:18:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

