

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042122\
 Data File : VX028253.D
 Acq On : 21 Apr 2022 19:52
 Operator : JC/MD
 Sample : N2511-06 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1530

Quant Time: Apr 22 06:19:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	387015	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	674962	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	634867	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	305250	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	278224	55.135	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	110.280%
35) Dibromofluoromethane	5.391	113	254245	53.712	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	107.420%
50) Toluene-d8	8.653	98	925678	50.917	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.840%
62) 4-Bromofluorobenzene	11.079	95	339571	48.849	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.700%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	5422	2.134	ug/l	95
20) Methylene Chloride	2.788	84	1289	0.344	ug/l #	87
25) 2-Butanone	4.568	43	8098	2.710	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042122\
Data File : VX028253.D
Acq On : 21 Apr 2022 19:52
Operator : JC/MD
Sample : N2511-06 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1530

Quant Time: Apr 22 06:19:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

