

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042122\
 Data File : VX028255.D
 Acq On : 21 Apr 2022 20:39
 Operator : JC/MD
 Sample : N2511-08 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1551

Quant Time: Apr 22 06:20:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	399244	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	692133	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	654441	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	315120	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	326788	62.775	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	125.560%
35) Dibromofluoromethane	5.391	113	296611	61.108	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	122.220%
50) Toluene-d8	8.653	98	1088109	58.366	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	116.740%
62) 4-Bromofluorobenzene	11.079	95	397624	55.782	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	111.560%
Target Compounds						
						Qvalue
18) Methyl Acetate	2.709	43	1775	0.383	ug/l	91
20) Methylene Chloride	2.794	84	1074	0.278	ug/l #	68
49) 1,4-Dioxane	7.678	88	3015	23.478	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042122\
Data File : VX028255.D
Acq On : 21 Apr 2022 20:39
Operator : JC/MD
Sample : N2511-08 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1551

Quant Time: Apr 22 06:20:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

