

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042122\
 Data File : VX028257.D
 Acq On : 21 Apr 2022 21:25
 Operator : JC/MD
 Sample : N2511-11 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1592

Quant Time: Apr 22 06:21:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	384909	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	666681	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	629426	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	295042	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	275162	54.827	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	= 109.660%		
35) Dibromofluoromethane	5.385	113	250505	53.580	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	= 107.160%		
50) Toluene-d8	8.653	98	910092	50.681	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	= 101.360%		
62) 4-Bromofluorobenzene	11.079	95	326014	47.482	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	= 94.960%		
Target Compounds						
20) Methylene Chloride	2.788	84	976	0.262	ug/l	Qvalue # 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042122\
Data File : VX028257.D
Acq On : 21 Apr 2022 21:25
Operator : JC/MD
Sample : N2511-11 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1592

Quant Time: Apr 22 06:21:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

