

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042122\
 Data File : VX028261.D
 Acq On : 21 Apr 2022 23:44
 Operator : JC/MD
 Sample : VX0421MBL02
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0421MBL02

Quant Time: Apr 22 06:23:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	376676	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	647068	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	609261	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	283805	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	271825	55.346	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	110.700%
35) Dibromofluoromethane	5.391	113	248258	54.708	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	109.420%
50) Toluene-d8	8.653	98	900731	51.680	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.360%
62) 4-Bromofluorobenzene	11.085	95	324793	48.738	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042122\
Data File : VX028261.D
Acq On : 21 Apr 2022 23:44
Operator : JC/MD
Sample : VX0421MBL02
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0421MBL02

Quant Time: Apr 22 06:23:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

