

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042122\
 Data File : VX028269.D
 Acq On : 22 Apr 2022 02:50
 Operator : JC/MD
 Sample : N2480-09ME
 Misc : 4.47g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-16-1A-3-GRME

Quant Time: Apr 22 06:31:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	593572	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	932460	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	842823	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	478760	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	301219	38.920	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	77.840%
35) Dibromofluoromethane	5.391	113	281228	43.006	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	86.020%
50) Toluene-d8	8.653	98	1060789	42.236	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	84.480%
62) 4-Bromofluorobenzene	11.085	95	429519	44.726	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.460%
Target Compounds						
					Qvalue	
31) Cyclohexane	5.458	56	10136	1.075	ug/l #	96
39) Methylcyclohexane	7.373	83	161680	15.022	ug/l	97
73) Isopropylbenzene	10.963	105	525081	15.957	ug/l	98
78) n-propylbenzene	11.305	91	155567	4.263	ug/l	100
83) tert-Butylbenzene	11.719	119	92278	3.235	ug/l	86
85) sec-Butylbenzene	11.896	105	755763	21.962	ug/l	96
89) n-Butylbenzene	12.335	91	510526	20.943	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042122\
Data File : VX028269.D
Acq On : 22 Apr 2022 02:50
Operator : JC/MD
Sample : N2480-09ME
Misc : 4.47g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-16-1A-3-GRME

Quant Time: Apr 22 06:31:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

