

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042122\
 Data File : VX028258.D
 Acq On : 21 Apr 2022 21:48
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 22 09:13:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	0.323	0.381	-18.0	136	0.00
3 P	Chloromethane	0.375	0.389	-3.7	122	0.00
4 C	Vinyl Chloride	0.504	0.478	5.2#	108	0.00
5 T	Bromomethane	0.522	0.415	20.5	87	0.00
6 T	Chloroethane	0.461	0.345	25.2#	99	0.00
7 T	Trichlorofluoromethane	1.112	0.921	17.2	94	0.00
8 T	Diethyl Ether	0.382	0.322	15.7	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.405	0.442	-9.1	138	0.00
10 T	Methyl Iodide	0.509	0.537	-5.5	111	0.00
11 T	Tert butyl alcohol	0.135	0.114	15.6	100	0.01
12 CM	1,1-Dichloroethene	0.384	0.418	-8.9#	137	0.00
13 T	Acrolein	0.080	0.086	-7.5	101	0.00
14 T	Allyl chloride	0.645	0.581	9.9	100	0.00
15 T	Acrylonitrile	0.266	0.245	7.9	97	0.00
16 T	Acetone	0.215	0.201	6.5	112	0.00
17 T	Carbon Disulfide	1.035	0.942	9.0	104	0.00
18 T	Methyl Acetate	0.580	0.534	7.9	105	0.00
19 T	Methyl tert-butyl Ether	1.602	1.528	4.6	98	0.00
20 T	Methylene Chloride	0.484	0.489	-1.0	118	0.00
21 T	trans-1,2-Dichloroethene	0.523	0.488	6.7	100	0.00
22 T	Diisopropyl ether	1.469	1.325	9.8	96	0.00
23 T	Vinyl Acetate	1.283	1.165	9.2	94	0.00
24 P	1,1-Dichloroethane	0.867	0.819	5.5	106	0.00
25 T	2-Butanone	0.386	0.335	13.2	92	0.00
26 T	2,2-Dichloropropane	0.847	0.652	23.0	82	0.00
27 T	cis-1,2-Dichloroethene	0.623	0.591	5.1	102	0.00
28 T	Bromochloromethane	0.372	0.323	13.2	90	0.00
29 T	Tetrahydrofuran	0.252	0.215	14.7	92	0.00
30 C	Chloroform	1.041	0.949	8.8#	97	0.00
31 T	Cyclohexane	0.794	0.687	13.5	92	0.00
32 T	1,1,1-Trichloroethane	0.943	0.865	8.3	96	0.00
33 S	1,2-Dichloroethane-d4	0.652	0.565	13.3	92	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	101	0.00
35 S	Dibromofluoromethane	0.351	0.319	9.1	96	0.00
36 T	1,1-Dichloropropene	0.457	0.411	10.1	96	0.00
37 T	Ethyl Acetate	0.464	0.407	12.3	93	0.00
38 T	Carbon Tetrachloride	0.516	0.476	7.8	95	0.00
39 T	Methylcyclohexane	0.577	0.504	12.7	93	0.00
40 TM	Benzene	1.339	1.236	7.7	96	0.00
41 T	Methacrylonitrile	0.239	0.218	8.8	97	0.00
42 TM	1,2-Dichloroethane	0.497	0.437	12.1	93	0.00
43 T	Isopropyl Acetate	0.755	0.651	13.8	93	0.00
44 TM	Trichloroethene	0.388	0.369	4.9	101	0.00
45 C	1,2-Dichloropropane	0.328	0.294	10.4#	96	0.00
46 T	Dibromomethane	0.257	0.237	7.8	99	0.00
47 T	Bromodichloromethane	0.494	0.443	10.3	93	0.00
48 T	Methyl methacrylate	0.347	0.305	12.1	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042122\
 Data File : VX028258.D
 Acq On : 21 Apr 2022 21:48
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 22 09:13:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	96	0.00
50 S	Toluene-d8	1.347	1.175	12.8	91	0.00
51 T	4-Methyl-2-Pentanone	0.485	0.422	13.0	92	0.00
52 CM	Toluene	0.930	0.850	8.6#	95	0.00
53 T	t-1,3-Dichloropropene	0.525	0.461	12.2	88	0.00
54 T	cis-1,3-Dichloropropene	0.564	0.489	13.3	89	0.00
55 T	1,1,2-Trichloroethane	0.378	0.355	6.1	99	0.00
56 T	Ethyl methacrylate	0.548	0.504	8.0	94	0.00
57 T	1,3-Dichloropropane	0.610	0.554	9.2	96	0.00
58 T	2-Chloroethyl Vinyl ether	0.257	0.239	7.0	92	0.00
59 T	2-Hexanone	0.375	0.332	11.5	93	0.00
60 T	Dibromochloromethane	0.412	0.378	8.3	92	0.00
61 T	1,2-Dibromoethane	0.415	0.389	6.3	95	0.00
62 S	4-Bromofluorobenzene	0.515	0.444	13.8	88	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	98	0.00
64 T	Tetrachloroethene	0.391	0.365	6.6	99	0.00
65 PM	Chlorobenzene	1.066	0.990	7.1	95	0.00
66 T	1,1,1,2-Tetrachloroethane	0.397	0.378	4.8	97	0.00
67 C	Ethyl Benzene	1.788	1.703	4.8#	94	0.00
68 T	m/p-Xylenes	0.728	0.699	4.0	94	0.00
69 T	o-Xylene	0.725	0.690	4.8	94	0.00
70 T	Styrene	1.175	1.147	2.4	94	0.00
71 P	Bromoform	0.313	0.292	6.7	91	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
73 T	Isopropylbenzene	3.437	3.119	9.3	93	0.00
74 T	N-amyl acetate	1.188	1.067	10.2	92	0.00
75 P	1,1,2,2-Tetrachloroethane	1.115	1.024	8.2	96	0.00
76 T	1,2,3-Trichloropropane	0.984	0.891	9.5	91	0.00
77 T	Bromobenzene	0.895	0.815	8.9	95	0.00
78 T	n-propylbenzene	3.811	3.529	7.4	92	0.00
79 T	2-Chlorotoluene	2.369	2.126	10.3	92	0.00
80 T	1,3,5-Trimethylbenzene	2.917	2.729	6.4	93	0.00
81 T	trans-1,4-Dichloro-2-butene	0.331	0.267	19.3	79	0.00
82 T	4-Chlorotoluene	2.729	2.471	9.5	91	0.00
83 T	tert-Butylbenzene	2.979	2.779	6.7	94	0.00
84 T	1,2,4-Trimethylbenzene	2.900	2.706	6.7	91	0.00
85 T	sec-Butylbenzene	3.594	3.356	6.6	93	0.00
86 T	p-Isopropyltoluene	3.100	2.935	5.3	92	0.00
87 T	1,3-Dichlorobenzene	1.682	1.591	5.4	94	0.00
88 T	1,4-Dichlorobenzene	1.766	1.631	7.6	94	0.00
89 T	n-Butylbenzene	2.546	2.350	7.7	90	0.00
90 T	Hexachloroethane	0.498	0.464	6.8	87	0.00
91 T	1,2-Dichlorobenzene	1.671	1.615	3.4	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.248	0.227	8.5	92	0.00
93 T	1,2,4-Trichlorobenzene	0.994	1.005	-1.1	103	0.00
94 T	Hexachlorobutadiene	0.443	0.402	9.3	96	0.00
95 T	Naphthalene	3.310	3.430	-3.6	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042122\
Data File : VX028258.D
Acq On : 21 Apr 2022 21:48
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 22 09:13:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.975	0.996	-2.2	102	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6