



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042122\
Data File : VX028230.D
Acq On : 21 Apr 2022 10:32
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Title : SW846 8260
Last Update : Tue Apr 19 13:38:04 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.8	10820	PASS
75	95	30	60	52.7	32008	PASS
95	95	100	100	100.0	60733	PASS
96	95	5	9	6.7	4074	PASS
173	174	0.00	2	1.4	649	PASS
174	95	50	100	78.1	47445	PASS
175	174	5	9	8.2	3881	PASS
176	174	95	101	95.6	45355	PASS
177	176	5	9	6.9	3111	PASS