



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042122\
Data File : VX028259.D
Acq On : 21 Apr 2022 22:34
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Title : SW846 8260
Last Update : Tue Apr 19 13:38:04 2022



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1632

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	17.0	11603	PASS
75	95	30	60	50.6	34459	PASS
95	95	100	100	100.0	68112	PASS
96	95	5	9	6.8	4638	PASS
173	174	0.00	2	1.2	679	PASS
174	95	50	100	86.4	58840	PASS
175	174	5	9	7.2	4258	PASS
176	174	95	101	97.3	57259	PASS
177	176	5	9	6.1	3520	PASS