

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042123\
 Data File : VX035233.D
 Acq On : 21 Apr 2023 16:03
 Operator : JC/MD
 Sample : 02429-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 230419054-04-TRIP-BLANK

Quant Time: Apr 22 01:06:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	249762	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	417165	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	375578	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	171722	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	152066	39.810	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	79.620%
35) Dibromofluoromethane	5.385	113	125552	45.503	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.000%
50) Toluene-d8	8.647	98	488260	47.942	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.880%
62) 4-Bromofluorobenzene	11.079	95	181937	46.164	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.320%
Target Compounds						
					Qvalue	
16) Acetone	2.398	43	29227	18.835	ug/l	96
30) Chloroform	5.092	83	4771	0.861	ug/l	96
47) Bromodichloromethane	7.830	83	1096	0.282	ug/l #	68
65) Chlorobenzene	10.079	112	2131	0.273	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042123\
Data File : VX035233.D
Acq On : 21 Apr 2023 16:03
Operator : JC/MD
Sample : 02429-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
230419054-04-TRIP-BLANK

Quant Time: Apr 22 01:06:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

