

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042125\
 Data File : VX045870.D
 Acq On : 21 Apr 2025 10:56
 Operator : JC/MD
 Sample : Q1841-03 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OILY-LIQUIDS

Quant Time: Apr 22 01:23:46 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	63683	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	127137	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	116045	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	48543	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	61476	52.787	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 105.580%	
35) Dibromofluoromethane	5.379	113	46070	51.073	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.140%	
50) Toluene-d8	8.647	98	156073	49.571	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 99.140%	
62) 4-Bromofluorobenzene	11.079	95	58394	50.918	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 101.840%	
Target Compounds						
16) Acetone	2.380	43	685	1.432	ug/l	Qvalue # 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042125\
Data File : VX045870.D
Acq On : 21 Apr 2025 10:56
Operator : JC/MD
Sample : Q1841-03 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OILY-LIQUIDS

Quant Time: Apr 22 01:23:46 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

