

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042125\
 Data File : VX045871.D
 Acq On : 21 Apr 2025 11:31
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Apr 22 01:23:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	66085	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	129472	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	120859	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	55062	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	63376	52.441	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.880%
35) Dibromofluoromethane	5.373	113	46190	50.283	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.560%
50) Toluene-d8	8.641	98	161148	50.260	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.520%
62) 4-Bromofluorobenzene	11.079	95	64291	55.049	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	110.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042125\
Data File : VX045871.D
Acq On : 21 Apr 2025 11:31
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Apr 22 01:23:55 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

