

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042125\
 Data File : VX045872.D
 Acq On : 21 Apr 2025 11:55
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Apr 22 01:24:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	67973	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	132652	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	121008	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49247	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64846	52.167	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.340%
35) Dibromofluoromethane	5.385	113	47000	49.938	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.880%
50) Toluene-d8	8.647	98	164131	49.963	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.920%
62) 4-Bromofluorobenzene	11.079	95	60704	50.731	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042125\
Data File : VX045872.D
Acq On : 21 Apr 2025 11:55
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Apr 22 01:24:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

