

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042219\
 Data File : VX009087.D
 Acq On : 22 Apr 2019 11:19
 Operator : JC/SP
 Sample : K2487-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB

Quant Time: Apr 23 03:28:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	250534	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	406946	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	352354	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	152123	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	164937	46.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.42%	
35) Dibromofluoromethane	5.50	113	123415	47.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.50%	
50) Toluene-d8	8.71	98	503784	48.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.74%	
62) 4-Bromofluorobenzene	11.13	95	169592	46.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.26%	
Target Compounds						
16) Acetone	2.44	43	6669	3.642	ug/l	93
20) Methylene Chloride	2.86	84	4765	1.522	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042219\
Data File : VX009087.D
Acq On : 22 Apr 2019 11:19
Operator : JC/SP
Sample : K2487-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB

Quant Time: Apr 23 03:28:36 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

