

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042219\
 Data File : VX009092.D
 Acq On : 22 Apr 2019 13:15
 Operator : JC/SP
 Sample : K2488-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Quant Time: Apr 23 03:41:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	246833	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	399923	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	349216	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	148794	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	162502	46.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.44%	
35) Dibromofluoromethane	5.50	113	121297	47.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.52%	
50) Toluene-d8	8.71	98	496315	48.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.98%	
62) 4-Bromofluorobenzene	11.13	95	163679	45.30	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.60%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042219\
Data File : VX009092.D
Acq On : 22 Apr 2019 13:15
Operator : JC/SP
Sample : K2488-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Quant Time: Apr 23 03:41:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

