

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042219\
 Data File : VX009100.D
 Acq On : 22 Apr 2019 16:22
 Operator : JC/SP
 Sample : K2515-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-12

Quant Time: Apr 23 04:07:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	239795	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	392198	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	333410	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	138249	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	159296	47.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.28%	
35) Dibromofluoromethane	5.50	113	119387	47.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.86%	
50) Toluene-d8	8.71	98	488702	48.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.36%	
62) 4-Bromofluorobenzene	11.13	95	159056	44.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.78%	
Target Compounds						
16) Acetone	2.44	43	10928	6.236	ug/l	98
18) Methyl Acetate	2.78	43	6583	1.505	ug/l	93
20) Methylene Chloride	2.85	84	209449	69.888	ug/l	97
43) Isopropyl Acetate	6.45	43	16414	1.951	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042219\
Data File : VX009100.D
Acq On : 22 Apr 2019 16:22
Operator : JC/SP
Sample : K2515-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-12

Quant Time: Apr 23 04:07:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

