

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042219\
 Data File : VX009102.D
 Acq On : 22 Apr 2019 17:09
 Operator : JC/SP
 Sample : IBLK
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Apr 23 04:13:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	245175	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	396390	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	349888	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	156659	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	160626	46.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.98%	
35) Dibromofluoromethane	5.50	113	119297	46.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.78%	
50) Toluene-d8	8.71	98	493930	48.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.38%	
62) 4-Bromofluorobenzene	11.13	95	169800	47.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.82%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042219\
Data File : VX009102.D
Acq On : 22 Apr 2019 17:09
Operator : JC/SP
Sample : IBLK
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Apr 23 04:13:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

