

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042219\
 Data File : VX009103.D
 Acq On : 22 Apr 2019 17:32
 Operator : JC/SP
 Sample : K2516-01DL 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P041919DL

Quant Time: Apr 23 04:21:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	249597	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	399502	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	348899	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	151065	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	163907	46.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.20%	
35) Dibromofluoromethane	5.50	113	120859	47.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.28%	
50) Toluene-d8	8.71	98	503452	49.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.48%	
62) 4-Bromofluorobenzene	11.13	95	167316	46.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.72%	
Target Compounds						
16) Acetone	2.44	43	134154	73.545	ug/l	99
25) 2-Butanone	4.67	43	229426	78.686	ug/l	99
29) Tetrahydrofuran	5.13	42	102733	52.642	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042219\
Data File : VX009103.D
Acq On : 22 Apr 2019 17:32
Operator : JC/SP
Sample : K2516-01DL 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P041919DL

Quant Time: Apr 23 04:21:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

