

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042219\
 Data File : VX009104.D
 Acq On : 22 Apr 2019 17:55
 Operator : JC/SP
 Sample : K2522-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WELL-1

Quant Time: Apr 23 04:25:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	235425	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	382551	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	328063	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	133992	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	157309	47.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.82%	
35) Dibromofluoromethane	5.50	113	116157	47.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.62%	
50) Toluene-d8	8.71	98	478966	48.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.84%	
62) 4-Bromofluorobenzene	11.13	95	154873	44.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.62%	
Target Compounds						
16) Acetone	2.45	43	3592	2.088	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042219\
Data File : VX009104.D
Acq On : 22 Apr 2019 17:55
Operator : JC/SP
Sample : K2522-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WELL-1

Quant Time: Apr 23 04:25:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

