

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042219\
 Data File : VX009094.D
 Acq On : 22 Apr 2019 14:02
 Operator : JC/SP
 Sample : K2449-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3-TB10-190415

Quant Time: Apr 23 03:47:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	245022	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	399159	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	335638	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	136181	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	163579	47.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.74%	
35) Dibromofluoromethane	5.50	113	120423	47.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.02%	
50) Toluene-d8	8.71	98	490538	48.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.04%	
62) 4-Bromofluorobenzene	11.13	95	157238	43.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.20%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042219\
Data File : VX009094.D
Acq On : 22 Apr 2019 14:02
Operator : JC/SP
Sample : K2449-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3-TB10-190415

Quant Time: Apr 23 03:47:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.66 min Scan# 740

Delta R.T. -0.00 min

Lab File: VX009094.D

Acq: 22 Apr 2019 14:02

Instrument :

MSVOA_X

ClientSampleId :


Z3-TB10-190415

Tot Ion:168 Resp: 245022

Ion Ratio Lower Upper

168 100

99 54.5 43.0 64.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.66

80000

60000



40000

20000

0

Time-->

5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 47.372 ug/l

RT: 6.06 min Scan# 805

Delta R.T. -0.00 min

Lab File: VX009094.D


Acq: 22 Apr 2019 14:02

Tgt Ion: 65 Resp: 163579

Ion Ratio Lower Upper

65 100

67 54.6 0.0 108.6



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.06

60000


40000


20000

0

Time-->

5.90 6.00 6.10 6.20 6.30









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.86 min Scan# 936
 Delta R.T. -0.00 min
 Lab File: VX009094.D
 Acq: 22 Apr 2019 14:02

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3-TB10-190415

Tot Ion	Ratio	Lower	Upper
114	100		
63	23.1	0.0	46.0
88	16.6	0.0	32.6




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VX0
 Ion 87.90 (87.60 to 88.60): VX0






#35
 Dibromofluoromethane
 Concen: 47.009 ug/l
 RT: 5.50 min Scan# 713
 Delta R.T. -0.00 min
 Lab File: VX009094.D
 Acq: 22 Apr 2019 14:02

Tgt Ion	Ratio	Lower	Upper
113	100		
111	103.1	82.9	124.3
192	21.5	17.4	26.0



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 48.016 ug/l
RT: 8.71 min Scan# 1240
Delta R.T. -0.00 min
Lab File: VX009094.D
Acq: 22 Apr 2019 14:02




Instrument :
MSVOA_X
ClientSampleId :
Z3-TB10-190415


Tot Ion: 98 Resp: 490538
Ion Ratio Lower Upper
98 100
100 64.6 51.4 77.2



#62
4-Bromofluorobenzene
Concen: 43.600 ug/l
RT: 11.13 min Scan# 1637
Delta R.T. -0.00 min
Lab File: VX009094.D
Acq: 22 Apr 2019 14:02

Tgt Ion: 95 Resp: 157238
Ion Ratio Lower Upper
95 100
174 79.8 0.0 159.6
176 77.2 0.0 154.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. -0.00 min
Lab File: VX009094.D
Acq: 22 Apr 2019 14:02

Instrument :
MSVOA_X
ClientSampleId :
Z3-TB10-190415

Tot Ion:117 Resp: 335638
Ion Ratio Lower Upper
117 100
82 60.7 48.2 72.4
119 32.6 25.0 37.6




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. -0.00 min
Lab File: VX009094.D
Acq: 22 Apr 2019 14:02

Tgt Ion:152 Resp: 136181
Ion Ratio Lower Upper
152 100
115 58.9 41.4 124.2
150 156.7 0.0 349.8



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

