

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042220\
 Data File : VX015846.D
 Acq On : 22 Apr 2020 17:01
 Operator : JC/SP
 Sample : L2363-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY038PS037-200420

Quant Time: Apr 23 12:11:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042220W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 23 04:36:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	909416	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1393346	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1353056	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	745692	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	506861	49.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.00%	
35) Dibromofluoromethane	5.47	113	409098	50.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.74%	
50) Toluene-d8	8.70	98	1658074	51.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.42%	
62) 4-Bromofluorobenzene	11.12	95	701329	52.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.60%	
Target Compounds						
16) Acetone	2.42	43	124815	27.714	ug/l	99
25) 2-Butanone	4.65	43	20306	2.853	ug/l	94
27) cis-1,2-Dichloroethene	4.56	96	10413	1.038	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042220\
Data File : VX015846.D
Acq On : 22 Apr 2020 17:01
Operator : JC/SP
Sample : L2363-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY038PS037-200420

Quant Time: Apr 23 12:11:40 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042220W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 23 04:36:12 2020
Response via : Initial Calibration

