

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042220\
 Data File : VX015849.D
 Acq On : 22 Apr 2020 18:09
 Operator : JC/SP
 Sample : L2379-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS037MW162-200420

Quant Time: Apr 23 12:12:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042220W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 23 04:36:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	892197	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1382494	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1398441	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	750947	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.04	65	508837	50.14	ug/l	0.00
Spiked Amount			Recovery	=	100.28%	
35) Dibromofluoromethane	5.47	113	402887	50.49	ug/l	0.00
Spiked Amount			Recovery	=	100.98%	
50) Toluene-d8	8.70	98	1697820	52.85	ug/l	0.00
Spiked Amount			Recovery	=	105.70%	
62) 4-Bromofluorobenzene	11.12	95	727146	55.18	ug/l	0.00
Spiked Amount			Recovery	=	110.36%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.39	62	80857	8.931	ug/l	100
12) 1,1-Dichloroethene	2.36	96	11129	1.424	ug/l	92
21) trans-1,2-Dichloroethene	3.15	96	11704	1.350	ug/l	92
24) 1,1-Dichloroethane	3.67	63	18659	1.196	ug/l #	93
27) cis-1,2-Dichloroethene	4.57	96	5288080	537.219	ug/l	89
32) 1,1,1-Trichloroethane	5.47	97	9124	0.652	ug/l #	35
44) Trichloroethene	7.19	130	7348	0.566	ug/l	85
64) Tetrachloroethene	9.32	164	31907	1.693	ug/l	93
83) tert-Butylbenzene	11.75	119	35072	0.972	ug/l	100
85) sec-Butylbenzene	11.93	105	19062	0.401	ug/l #	5

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042220\
Data File : VX015849.D
Acq On : 22 Apr 2020 18:09
Operator : JC/SP
Sample : L2379-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS037MW162-200420

Quant Time: Apr 23 12:12:53 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042220W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 23 04:36:12 2020
Response via : Initial Calibration

