

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042220\
 Data File : VX015850.D
 Acq On : 22 Apr 2020 18:31
 Operator : JC/SP
 Sample : L2379-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS037MW012-200420

Quant Time: Apr 23 12:13:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042220W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 23 04:36:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	925882	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1444491	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1438826	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	795954	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	533644	50.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.36%	
35) Dibromofluoromethane	5.47	113	421750	50.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.16%	
50) Toluene-d8	8.70	98	1746642	52.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.08%	
62) 4-Bromofluorobenzene	11.12	95	730852	53.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.16%	
Target Compounds						
4) Vinyl Chloride	1.39	62	10506	1.118	ug/l #	87
27) cis-1,2-Dichloroethene	4.57	96	25921	2.538	ug/l	84
83) tert-Butylbenzene	11.76	119	56939	1.489	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042220\
Data File : VX015850.D
Acq On : 22 Apr 2020 18:31
Operator : JC/SP
Sample : L2379-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS037MW012-200420

Quant Time: Apr 23 12:13:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042220W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 23 04:36:12 2020
Response via : Initial Calibration

