

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042220\
 Data File : VX015851.D
 Acq On : 22 Apr 2020 18:54
 Operator : JC/SP
 Sample : L2379-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST006MW148-200420

Manual Integrations
 APPROVED

MMDadoda
 4/23/2020 12:15:04 PM

Quant Time: Apr 23 12:13:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042220W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 23 04:36:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	952457	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1438974	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1470692	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	785827	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	524023	48.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.74%	
35) Dibromofluoromethane	5.47	113	422032	50.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.62%	
50) Toluene-d8	8.70	98	1804842	53.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.96%	
62) 4-Bromofluorobenzene	11.12	95	752770	54.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.76%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.39	62	392030	40.563	ug/l	99
16) Acetone	2.42	43	42924	9.100	ug/l	97
24) 1,1-Dichloroethane	3.67	63	20765	1.247	ug/l	99
27) cis-1,2-Dichloroethene	4.57	96	387645	36.890	ug/l	88
31) Cyclohexane	5.55	56	34746	2.195	ug/l #	86
32) 1,1,1-Trichloroethane	5.47	97	21568	1.444	ug/l #	47
40) Benzene	6.12	78	29374	0.765	ug/l	97
65) Chlorobenzene	10.12	112	17055	0.580	ug/l	96
67) Ethyl Benzene	10.24	91	16770	0.325	ug/l	90
73) Isopropylbenzene	11.00	105	19131	0.383	ug/l	85
83) tert-Butylbenzene	11.76	119	61170	1.620	ug/l	91
85) sec-Butylbenzene	11.93	105	175801	3.533	ug/l #	56
89) n-Butylbenzene	12.37	91	11524m	0.277	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042220\
Data File : VX015851.D
Acq On : 22 Apr 2020 18:54
Operator : JC/SP
Sample : L2379-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST006MW148-200420

Manual Integrations
APPROVED

MMDadoda
4/23/2020 12:15:04 PM

Quant Time: Apr 23 12:13:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042220W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 23 04:36:12 2020
Response via : Initial Calibration

