

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042220\
 Data File : VX015852.D
 Acq On : 22 Apr 2020 19:17
 Operator : JC/SP
 Sample : L2379-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS037TW143-200420

Quant Time: Apr 23 12:14:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042220W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 23 04:36:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	883749	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1386198	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1345452	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	734246	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	503559	50.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.20%	
35) Dibromofluoromethane	5.47	113	401371	50.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.32%	
50) Toluene-d8	8.70	98	1647144	51.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.28%	
62) 4-Bromofluorobenzene	11.12	95	698313	52.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.70%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.57	96	97139	9.963	ug/l	90
44) Trichloroethene	7.18	130	16403	1.261	ug/l	71
64) Tetrachloroethene	9.32	164	31186	1.720	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042220\
Data File : VX015852.D
Acq On : 22 Apr 2020 19:17
Operator : JC/SP
Sample : L2379-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS037TW143-200420

Quant Time: Apr 23 12:14:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042220W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 23 04:36:12 2020
Response via : Initial Calibration

