

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042220\
 Data File : VX015853.D
 Acq On : 22 Apr 2020 19:39
 Operator : JC/SP
 Sample : L2379-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST008RW258-200420

Quant Time: Apr 23 12:14:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042220W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 23 04:36:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	909483	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1402098	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1372459	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	758054	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	514517	49.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.48%	
35) Dibromofluoromethane	5.47	113	408199	50.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.88%	
50) Toluene-d8	8.70	98	1700834	52.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.40%	
62) 4-Bromofluorobenzene	11.12	95	714690	53.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.94%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.37	101	16333	2.103	ug/l	98
16) Acetone	2.42	43	23825	5.290	ug/l	95
27) cis-1,2-Dichloroethene	4.57	96	271603	27.068	ug/l	89
44) Trichloroethene	7.19	130	135009	10.262	ug/l	94
64) Tetrachloroethene	9.32	164	385934	20.862	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042220\
Data File : VX015853.D
Acq On : 22 Apr 2020 19:39
Operator : JC/SP
Sample : L2379-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST008RW258-200420

Quant Time: Apr 23 12:14:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042220W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 23 04:36:12 2020
Response via : Initial Calibration

