

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042220\
 Data File : VX015854.D
 Acq On : 22 Apr 2020 20:02
 Operator : JC/SP
 Sample : L2379-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST006MW015-200420

Quant Time: Apr 23 12:15:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042220W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 23 04:36:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	952719	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1444283	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1393353	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	774145	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	530656	48.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.94%	
35) Dibromofluoromethane	5.47	113	420455	50.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.86%	
50) Toluene-d8	8.70	98	1738277	51.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.60%	
62) 4-Bromofluorobenzene	11.12	95	718570	52.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.38%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.57	96	436084	41.488	ug/l	90
44) Trichloroethene	7.20	130	45589	3.364	ug/l	92
64) Tetrachloroethene	9.32	164	82299	4.382	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042220\
Data File : VX015854.D
Acq On : 22 Apr 2020 20:02
Operator : JC/SP
Sample : L2379-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST006MW015-200420

Quant Time: Apr 23 12:15:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042220W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 23 04:36:12 2020
Response via : Initial Calibration

