

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042220\
 Data File : VX015862.D
 Acq On : 23 Apr 2020 00:33
 Operator : JC/SP
 Sample : L2380-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY001MW003-200421

Quant Time: Apr 23 12:22:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042220W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 23 04:36:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	916548	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1411493	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1388139	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	748349	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.04	65	510782	49.00	ug/l	0.00
Spiked Amount			Recovery	=	98.00%	
35) Dibromofluoromethane	5.47	113	409067	50.21	ug/l	0.00
Spiked Amount			Recovery	=	100.42%	
50) Toluene-d8	8.70	98	1718562	52.40	ug/l	0.00
Spiked Amount			Recovery	=	104.80%	
62) 4-Bromofluorobenzene	11.12	95	710474	52.80	ug/l	0.00
Spiked Amount			Recovery	=	105.60%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	3.01	59	681267	319.860	ug/l	99
16) Acetone	2.42	43	15394	3.391	ug/l #	83
31) Cyclohexane	5.56	56	25392	1.667	ug/l #	84
39) Methylcyclohexane	7.45	83	59257	3.740	ug/l	94
73) Isopropylbenzene	11.00	105	58227	1.225	ug/l	93
78) n-propylbenzene	11.34	91	59025	1.084	ug/l	100
83) tert-Butylbenzene	11.76	119	10001	0.278	ug/l	87
85) sec-Butylbenzene	11.93	105	20392	0.430	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042220\
Data File : VX015862.D
Acq On : 23 Apr 2020 00:33
Operator : JC/SP
Sample : L2380-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY001MW003-200421

Quant Time: Apr 23 12:22:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042220W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 23 04:36:12 2020
Response via : Initial Calibration

