

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042221\
 Data File : VX021813.D
 Acq On : 22 Apr 2021 13:55
 Operator : JC/MD
 Sample : M2114-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-02

Quant Time: Apr 23 07:42:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 20 13:46:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	150544	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	267788	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	282585	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	145513	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	85538	50.09	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 100.18%		
35) Dibromofluoromethane	5.464	113	86334	50.41	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.82%		
50) Toluene-d8	8.696	98	327765	49.78	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 99.56%		
62) 4-Bromofluorobenzene	11.116	95	139352	56.67	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 113.34%		
Target Compounds						
16) Acetone	2.416	43	6441	10.97	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042221\
Data File : VX021813.D
Acq On : 22 Apr 2021 13:55
Operator : JC/MD
Sample : M2114-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB-02

Quant Time: Apr 23 07:42:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 20 13:46:22 2021
Response via : Initial Calibration

