

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042221\
 Data File : VX021815.D
 Acq On : 22 Apr 2021 14:40
 Operator : JC/MD
 Sample : M2134-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 210421003-05-TRIP-BLANK

Quant Time: Apr 23 07:42:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 20 13:46:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	167828	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	300304	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	317166	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	169753	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	95809	50.32	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.64%	
35) Dibromofluoromethane	5.464	113	96819	50.41	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.82%	
50) Toluene-d8	8.695	98	370816	50.22	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.44%	
62) 4-Bromofluorobenzene	11.116	95	157852	57.24	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 114.48%	
Target Compounds						Qvalue
16) Acetone	2.422	43	3186	4.87	ug/l	93
30) Chloroform	5.178	83	1690	0.53	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042221\
Data File : VX021815.D
Acq On : 22 Apr 2021 14:40
Operator : JC/MD
Sample : M2134-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
210421003-05-TRIP-BLANK

Quant Time: Apr 23 07:42:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 20 13:46:22 2021
Response via : Initial Calibration

