

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042221\
 Data File : VX021817.D
 Acq On : 22 Apr 2021 15:26
 Operator : JC/MD
 Sample : M2114-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GMW-7

Quant Time: Apr 23 07:43:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 20 13:46:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	156388	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	281898	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	296042	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	152786	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	90265	50.88	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 101.76%	
35) Dibromofluoromethane	5.464	113	90661	50.29	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.58%	
50) Toluene-d8	8.695	98	345720	49.88	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.76%	
62) 4-Bromofluorobenzene	11.116	95	146847	56.73	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 113.46%	
Target Compounds						
16) Acetone	2.416	43	2881	4.72	ug/l	97
20) Methylene Chloride	2.831	84	985	0.54	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042221\
Data File : VX021817.D
Acq On : 22 Apr 2021 15:26
Operator : JC/MD
Sample : M2114-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GMW-7

Quant Time: Apr 23 07:43:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 20 13:46:22 2021
Response via : Initial Calibration

