

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042221\
 Data File : VX021822.D
 Acq On : 22 Apr 2021 17:19
 Operator : JC/MD
 Sample : M2138-22
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-FB-042021

Quant Time: Apr 23 07:44:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 20 13:46:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	154850	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	280659	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	297820	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	159128	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	88522	50.39	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.78%
35) Dibromofluoromethane	5.464	113	89190	49.69	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.38%
50) Toluene-d8	8.696	98	343449	49.77	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.54%
62) 4-Bromofluorobenzene	11.116	95	148851	57.76	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	115.52%
Target Compounds						
16) Acetone	2.422	43	1712	2.83	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042221\
Data File : VX021822.D
Acq On : 22 Apr 2021 17:19
Operator : JC/MD
Sample : M2138-22
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-FB-042021

Quant Time: Apr 23 07:44:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 20 13:46:22 2021
Response via : Initial Calibration

