

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042221\
 Data File : VX021824.D
 Acq On : 22 Apr 2021 18:05
 Operator : JC/MD
 Sample : M2138-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-MW-5S

Quant Time: Apr 23 07:45:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 20 13:46:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	152537	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	275254	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	287877	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	152344	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	87826	50.75	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.50%
35) Dibromofluoromethane	5.464	113	88031	50.01	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.02%
50) Toluene-d8	8.695	98	334991	49.50	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.00%
62) 4-Bromofluorobenzene	11.116	95	145096	57.41	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	114.82%
Target Compounds						
					Qvalue	
16) Acetone	2.422	43	2911	4.89	ug/l	97
64) Tetrachloroethene	9.317	164	1883	0.84	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042221\
Data File : VX021824.D
Acq On : 22 Apr 2021 18:05
Operator : JC/MD
Sample : M2138-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-MW-5S

Quant Time: Apr 23 07:45:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 20 13:46:22 2021
Response via : Initial Calibration

