

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042221\
 Data File : VX021825.D
 Acq On : 22 Apr 2021 18:27
 Operator : JC/MD
 Sample : M2138-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-DDC-2-PD

Manual Integrations
 APPROVED

MMDadoda
 4/26/2021 4:52:16 PM

Quant Time: Apr 23 07:45:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 20 13:46:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	154943	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	277307	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.091	117	297292	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	162956	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	89416	50.87	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 101.74%			
35) Dibromofluoromethane	5.464	113	90216	50.87	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.74%			
50) Toluene-d8	8.695	98	340963	50.00	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 100.00%			
62) 4-Bromofluorobenzene	11.122	95	149393	58.67	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 117.34%			
Target Compounds						Qvalue
30) Chloroform	5.172	83	2250	0.76	ug/l	99
38) Carbon Tetrachloride	5.751	117	3783	1.41	ug/l	86
44) Trichloroethene	7.190	130	2695	1.26	ug/l	81
64) Tetrachloroethene	9.317	164	1852459	796.91	ug/l	94
90) Hexachloroethane	12.579	117	39537m	20.59	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042221\
Data File : VX021825.D
Acq On : 22 Apr 2021 18:27
Operator : JC/MD
Sample : M2138-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-DDC-2-PD

Manual Integrations
APPROVED

MMDadoda
4/26/2021 4:52:16 PM

Quant Time: Apr 23 07:45:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 20 13:46:22 2021
Response via : Initial Calibration

