

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042221\
 Data File : VX021830.D
 Acq On : 22 Apr 2021 20:21
 Operator : JC/MD
 Sample : M2138-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-MW-1D(ONSITE)

Quant Time: Apr 23 07:46:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 20 13:46:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	157654	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	281743	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.097	117	301204	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.060	152	161200	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	91403	51.11	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.22%	
35) Dibromofluoromethane	5.464	113	92125	51.13	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.26%	
50) Toluene-d8	8.695	98	349138	50.40	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.80%	
62) 4-Bromofluorobenzene	11.122	95	150562	58.20	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 116.40%	
Target Compounds						
						Qvalue
16) Acetone	2.422	43	2450	3.98	ug/l	90
44) Trichloroethene	7.183	130	2904	1.34	ug/l	85
64) Tetrachloroethene	9.317	164	2790	1.18	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042221\
Data File : VX021830.D
Acq On : 22 Apr 2021 20:21
Operator : JC/MD
Sample : M2138-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-MW-1D(ONSITE)

Quant Time: Apr 23 07:46:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042021W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 20 13:46:22 2021
Response via : Initial Calibration

